

**INDIAN INSTITUTE OF TROPICAL METEOROLOGY,  
PASHAN, PUNE – 411 008, INDIA**

**File No.PS/125/27/2009**

**Technical Specifications for “SYSTEM OF AIR POLLUTION FORECASTING AND RESEARCH (SAFAR-2010)”**

The Manufacturer of Ambient Air Monitoring System should have world-wide recognition of having enough field experience in designing such system to provide high quality air pollution data. The System should be designed to minimize the time spent on maintenance and calibration. During establishing SAFAR system, following criteria /features must be adopted and the manufacturers must comply the same in order to provide high quality data.

1. Since the system is to be operated & used for data collection during the period of Commonwealth Games 2010 at Delhi, it is highly desirable that the manufacturer of the prime analysers (NO<sub>x</sub>, CO, O<sub>3</sub>, etc.) must have enough expertise on operation and maintenance of such monitoring system & on data reporting of the pollutants. Also, the preference will be given to the manufacturer who has ISO9000 /ISO9001 /ISO17025 Quality Accreditation /NATA accreditation to perform such operation & data reporting job. A certificate of an accredited authority to execute data reporting from ambient air and emission monitoring systems must be enclosed.
2. The CO, O<sub>3</sub> and NO<sub>x</sub> analysers, Gas calibrator, zero air generator, Data Acquisition Systems and Data Collection and Reporting software should preferably be manufactured by one manufacturer to provide substantial control over the quality of the system and to allow the design of the system to be best suited for application. The analysers used in the monitoring system must function efficiently in the temperature range of 5°C - 45°C as the system is to be operated on open roof tops or in mobile module during CWG-2010.
3. All analysers (CO, O<sub>3</sub>, NO<sub>x</sub>, PM10, PM2.5, Black carbon, Benzene) used in the system must be designed on US-EPA Equivalent Reference Methods. A certificate of US-EPA approval must be enclosed as qualifying requirement for bid.
4. All analysers in the system must have technical sophistication to be connected digitally to Data Acquisition System for providing substantial advantages in communication.
5. The Data Acquisition System must have features like automatic data validation software complete with data capture percentage.
6. Analysers for NO<sub>x</sub>, O<sub>3</sub> and CO should have a fold down front– to provide easy access to expendable items such as particulate filters and charcoal scrubbers, and power control switches
7. Analysers for the measurements of NO<sub>x</sub>, O<sub>3</sub> and CO should have both RS 232 and USB ports.

8. The Data Acquisition System should operate on Windows XP PRO Platform and be of a “Client/Server” configuration for improved reliability.
9. Real time or averaged data must be viewed quickly and easily through a remote CLIENT interface on the central computer
10. The Data Acquisition System feature calculated channels including nested calculations, vector averaging and rolling averages.
11. The System must support remote communication with full user interface from anywhere in the world via a TCP/IP network or dial up network.
12. Data Evaluation and Reporting Software must have been upgradeable to a multi-user version.
13. The Data Evaluation and Reporting Software must be able to collect and process data from a wide variety of data acquisition systems & analysers.
14. The engineer responsible for installation and commissioning of the station must provide on site training for operation, maintenance & data validation as per ISO9001/ISO17025 Quality Accreditation/ NATA accreditation procedure. This provides with hands-on experience and provides them with an understanding on how the system is integrated.
15. The manufacturer should provide the advice and services of his experts in preparing air pollution forecast if they have rendered such consultations and services in connection with yesteryears’ Commonwealth /Olympic Games or in connection with any such kind of prestigious events for air quality data interpretation. The experts must visit two times every six month and preferably all the days during 1-15<sup>th</sup> October 2010 (CWG-2010 window period) to ensure that Air quality monitoring system is performing well and providing useful data.

**Note: Supplier should categorically provide detail feedback on each and every point mentioned above (Total 15) in chronicle order related to compliance (yes) /non-compliance (no) or any other explanation/justification in a tabular form.**

<b>A. SPECIFICATIONS FOR OZONE ANALYSER -10 sets</b>		
1	General Range	User selectable and automatic
2	Concentración Range	0-1000 ppb
3	Lower Detectable Limit	≤0.5 ppb
4	Zero drift	< 1 ppb for 30 days
5	Response time	Programmable (10 s to 80 s)
6	Averaging time	Programmable (1 min to 24 hrs)
7	Internal memory	Suitable for data storage (more than 5 MB)
8	Data storage	Built in analyzer (should store 1 min average for more that 5 day)
9	Port & compatibility	RS-232 or RS-422 and Compatible with windows-XP
10	Analog output	0-10 V and 0-20 mA or 4-20 mA
11	Power supply	230 V AC - 50 Hz
12	Manual	Instruction & operation manuals
13	Calibration unit	In Built Calibration unit with internal zero span check feature
14	Approval	US-EPA approved

<b>B. SPECIFICATIONS FOR NO<sub>x</sub> -ANALYSER -10 sets</b>		
1	General Range	User selectable and automatic
2	Concentration Range	0-20 ppm
3	Lowest detectable limit	<0.4 ppb
4	Zero drift	<0.1 ppm per day
5	Response time	Programmable (30 s to 90 s)
6	Averaging time	Programmable (1 min to 24 hrs)
7	Internal memory	Suitable for data storage (more than 5 MB)
8	Data storage	Built in analyzer (should store 1 min average for more that 5 day)
9	Port & compatibility	RS-232 or RS-422 and Compatible with windows-XP
10	Analog output	0-10 V and 0-20 mA or 4-20 mA
11	Power supply	230 V AC - 50 Hz
12	Manual	Instruction & operation manuals
13	Calibration unit	Inbuilt unit Inbuilt unit for zero gas calibration and NO <sub>x</sub> ( preferable as permeation bench )
14	Approval	US-EPA approved

<b>C. SPECIFICATIONS FOR CO -ANALYSER (CO<sub>2</sub> as optional) -10 sets</b>		
1	General Range	User selectable and automatic
2	Concentration range	0-200 ppm
3	Lowest detectable limit	≤0.4 ppm
4	Zero drift	<0.1 ppm per day
5	Span Drift	1% of FS per day
6	Response time	Programmable (10 s to 80 s)
7	Averaging time	Programmable (1 min to 24 hrs)
8	Internal memory	Suitable for data storage (more than 5 MB)
9	Data storage	Built in analyzer (should store 1 min average for more that 5 day)
10	Port & compatibility	RS-232 or RS-422 and Compatible with windows-XP
11	Analog output	0-10 V and 0-20 mA or 4-20 mA
12	Power supply	230 V AC - 50 Hz
13	Manual	Instruction & operation manuals
13	Calibration unit	Inbuilt unit for zero gas calibration with solenoid valve for external span gas cylinder
14	Approval	US-EPA approved
15	<b>OPTIONAL</b> (Quote separately as additional cost)	<b>Specifications for CO<sub>2</sub> option:</b> Carbon dioxide measurement inbuilt in CO-analyser box Measurement range 50 to 2000 ppmv Accuracy (including repeatability, non-linearity and calibration uncertainty) ± (2 % of range + 2% of reading) Long-term stability <±5 % of range / 5 years Response time (63%) 1 minute

<b>D. SPECIFICATIONS FOR BENZENE ANALYSER -5 sets</b>		
1	General Range	0 - 100 ppb
2	Type	Gas Chromatograph with PID detector
3	Cycle	Min 10 min
4	Averaging time	Programmable (10 min to 24 hrs)
5	Internal memory	Suitable for data storage (more than one month)
6	Data storage	Built in analyzer (should store 10 min average for more that 5 day)
7	Port & compatibility	RS-232 or RS-422 and Compatible with windows-XP
8	Analog output	0 -10 V and 0-20 mA or 4-20 mA
9	Power supply	230 V AC - 50 Hz
10	Manual	Instruction & operation manuals
11	Operating software	Should be compatible with windows-XP
11	Calibration unit	Calibration switch for zero and span calibration unit with auto calibration or separately
12	Approval	US-EPA approved

<b>E. SPECIFICATIONS FOR SPM ANALYSER -10 sets</b>		
1	General Range	PM-10/PM-2.5 with Mass Distribution (online monitor)
2	Concentration Range	0.1 $\mu\text{g}/\text{m}^3$ - 10 $\text{mg}/\text{m}^3$
3	Lowest detectable limit	1 $\mu\text{g}/\text{m}^3$
4	Averaging time	Programmable (1 min to 24 hrs)
5	Internal memory	Suitable for data storage (more than 5MB)
6	Data storage	Built in analyzer (should store 1 min average for more that 5 day)
7	Port & compatibility	RS-232 or RS-422 and Compatible with windows-XP
8	Analog output	0 -10 V and 0-20 mA or 4-20 mA
9	Power supply	230 V AC - 50 Hz
10	Manual	Instruction & operation manuals
11	Calibration unit	Inbuilt Calibration unit with auto calibration or separately
12	Approval	US-EPA approved
13	Sample inlet	Each set should contain inlets/filter for PM-10 as well as PM-2.5

<b>F. SPECIFICATIONS FOR BLACK CARBON ANALYSER -5 sets</b>		
1	General Range	0.5 $\mu\text{g}/\text{m}^3$ - 10 $\text{mg}/\text{m}^3$
2	Type	Optical transmission absorption
3	Cycle	10 sec 1hour
4	Sensivity	1 $\text{ng}/\text{m}^3$ or less
4	Averaging time	Programmable (1 min to 24 hrs)
5	Internal memory	Suitable for data storage (more than one month)
6	Data storage	Built in analyzer (should store 10 min average for more that 5 day)
7	Port & compatibility	RS-232 or RS-422 and Compatible with windows-XP
8	Analog output	0 -10 V and 0-20 mA or 4-20 mA
9	Power supply	230 V AC - 50 Hz
10	Manual	Instruction & operation manuals
11	Operating software	Should compatible with windows-XP
12	Calibration unit	Calibration switch for zero and span calibration unit with auto calibration or separately
13	Approval	US-EPA approved
14	Vibration	Stabilization for vibration

<b>G. SPECIFICATIONS FOR ZERO GAS GENERATOR -10 sets</b>		
1	Outflow Generator	1 TO 5 Liter/Minute
2	Outlet Pressure	Adjustable Between 0.1 and 6 Bar
3	Purity	>99.999%
4	Power Supply	230 V, 50Hz
5	Manual	Instruction and Operation Manual
<b>Scrubbers</b> Hopcalite Purafil Charcoal Molecular Sieve Power Pressure Flow		Removal of CO (<0.1 ppb) Removal of NO, NO <sub>2</sub> , NO <sub>x</sub> (<0.1 ppb) Removal of SO <sub>2</sub> , H <sub>2</sub> S, Ozone (<0.1 ppb) Removal of H <sub>2</sub> O (Dew point -15 C) 220 volts AC, 50 Hz (optional 110VAC) Up to 200 KPA 0-9 LMP

<b>H. SPECIFICATIONS FOR MULTI POINT GAS CALIBRATOR (MGC) -10 sets</b>		
1	Power Supply	230 V; 50 Hz
2	Weight	< 20 Kg
3	Pressure	Ajustable Between 0.1 and 6 Bar
4	Dilution	From High Concentration Span Gas to Diluted Gas
5	Manual	Instruction and Operation Manual
6	Calibration Gases, Cylinders & Regulator	Cylinder of Capacity 10 liter with SS Regulator and single cylinder for all the calibration gases. All the calibration gases must be supplied in the one cylinder.

<b>Sr. No.</b>	<b>I. SPECIFICATIONS FOR LOCAL AND CENTRAL DATA ACQUISITION, CONTROL SYSTEM &amp; GSM-NETWORK</b>
1	Data logger system capable to acquire all the local station data.
2	GSM networking (or alternative) with minimum memory to store all the local station data for the period of 6 months, with capability of acquiring real time and store data from local station to centralization by GSM or other suitable network with software for the data analysis and visualization.
<b>LOCAL STATION: 10</b> <ol style="list-style-type: none"> <li>Should connect through RS232 and USB communication ports.</li> <li>Should able to transmit data and /remote control via modem/GSM network.</li> <li>Should provide GUI windows based application software for continuous air monitoring.</li> <li>Should provide calibration windows for analyzer calibration from computer.</li> </ol>	

5. Main window for real time display major parameters & status of all analyzers & sensors.
6. Control panel window for control of each analyzer.
7. Real time status and diagnostics for maintenance people.
8. Programmable down loading of data.
9. Should provide data storage capacity for at least 3 years.

**CENTRAL STATION: 1**

1. The software should have all advance features to enable remote automatic data collection from each station and should provide full control over data loggers from anywhere including real time viewing, status and diagnostic information.
2. Data communication and acquisition system should handle the data transmission /reception from 10 different monitoring stations of an ambient air quality network and receive incoming messages / signals from remote site. **In addition** to this it should also provide the similar connectivity and data control and transmission of 10 different automatic weather station data which are being installed there independently.
3. Central station software should manage 10 remote air quality-monitoring stations.
4. Should provide data management, analysis and reporting.
5. Should be compatible with Windows-XP
6. Should provide data storage capacity for at least 2 years.
7. The remote control ability on the client server model as such should look and realize exactly the same when used from a remote location as it does when used on site. It should be a separate process that is independent of what is being displayed on the logger.
8. Multiple clients should be able to connect to the server from different locations at a time to allow remote computer to view the data being logged in real time.

**Configuration of remote data acquisition machine-10**

Intel® Core™ 2 Quad CPU Q9400 @2.66 GHz processor (2.6 GHz or more)

3 GB DDR2 SDRAM

320 GB SATA Hard Drive

Integrated 7.1 Channel Audio

Genuine MS Windows XP professional licensed

16X DVD+/-RW Drive

Anti-virus software with 2 year licence

Integrated 7.1 Channel Audio

Integrated Intel® GMA X4500HD Graphics

17 inch flat TFT Panel monitor

Speakers with Headphone

Optical Mouse with pad and Key Board (Branded)

Ports: 2 front & 2 back side USB, 2 PS/2, 1 Serial, 1 Parallel Port

**Configuration of central data acquisition machine/workstation**

Same as above with 24 inches TFT flat monitor (instead of 17")

**Configuration of indoor digital data display system -2**

(One for central station Delhi and one for Model forecasting station Pune)

A Wall mounted LCD-Television of size 40 inches or above (LG-type or equivalent)

Should have HDMI (High-Definition Multimedia Interface) for PC and wise link USB 2.0 port for pendrive, etc.

Clear flat panel, high resolution (1024x768 or above), Natural true color, 3D ready

Dynamic Contrast ratio: 30,000:1 or more; Anti burn-in technology,

Movie plus, smart frame plus. Wide color enhancer

<b>J. SPECIFICATIONS FOR UPS /AIR CONDITIONER/RACK -10 sets</b>	
1	5 KVA Online Branded UPS With Minimum 4 Hour Battery Backup For all analyzer.
2	Branded Air Conditioner System split type of two ton (2 units) (hence total 20 units)
3	Standard mounting rack to accommodate all analyzers and necessary accessories (cylinders, calibrator, etc) with sample handling moisture removal manifold system, power distribution board, exhaust/cooling, dust filter, telescopic slides, etc

<b>K. OMC (OPERATIONAL AND MAINTENANCE CONTRACT):</b>
<p>Total Comprehensive Operational and Maintenance Contract (OMC) for 2 years for all set of analyzers including complete manned operation and supply of the data to central station. This should include all consumable; spare parts; regular maintenance and regular operation by deputing dedicated enough technical persons /engineers on permanent basis. These engineers will maintain analysers, manage all consumable, keep it up to date and carry out the complete operational analysis at all 10 different stations (spread around 50 km radius within Delhi and surrounding areas near CWG village) on routine basis. Quote separately for each year. Remember that for the first year the CMC will be free of cost and hence charge cost only for OMC-CMC. Instrument should remain operational at least 90% of the time in a year and should not remain non-functional continuously for more than 5 days in general and not more than 10 hours during most critical period Oct 1-15, 2010 during which CWG-2010 will be held. In case of non-compliance the OMC will be extended proportionately. The instrument should be commissioned, regularized and operational after the order is executed without any additional facility and cost to IITM excepting the infrastructure which includes well furnished laboratory space with 230V standard Indian power supply.</p> <p><b>Optional:</b> It may also be noted that after the completion of CWG-2010 (say from December 2010), all system including instruments may be shifted from Delhi to 10 different cities preferably be in central part of India for the remaining period of OMC. In addition, OMC may be extended for 3 more years after the expiry of present OMC. Hence vendor may quote additional cost for the same as optional item.</p>

## INDIAN INSTITUTE OF TROPICAL METEOROLOGY PUNE 411 008

### GENERAL TERMS & CONDITIONS

ENQUIRY NO: **PS/125/27/2009/**

- 1) The Tenderers are requested to give detailed sealed tender in their own forms in two Bids i.e.  
Part – I Technical Bid.  
Part - II Commercial Bid, both the bids addressed to the Director, Indian Institute of Tropical Meteorology, Dr. Homi Bhabha Road, NCL Post, Pashan, Pune – 411 008, INDIA.
- 2) This tender is not transferable.
- 3) a) The Equipments are required to be installed at **IITM, Pune** and subsequently Training is to be provided to the concerned scientists.  
b) Packing, Forwarding, Freight & Insurance and Commissioning Charges, if any extra may be quoted separately in Commercial Bid.  
c) In case if your quote is F.O.R./F.O.B. basis, estimated Insurance coverage charges may please be indicated.  
d) CIF, Mumbai value both by Airfreight.  
e) Octroi Payment if an, the same may please be shown separately.
- 5) Cost of the items should be mentioned clearly in the Commercial Offer (Part-II) only. The optional and any other essential items / accessories required for the maintenance of the equipment for the next three years should also be specified in the offer separately.
- 6) In case of foreign quote, the Principal supplier should clearly indicate the address of the Indian Agent and percentage (%) of Agency Commission payable if any, to be paid to the Indian Agent in Indian Currency.
- 7) Percentage & amount of Indian Agency commission payable in Indian currency after the receipt of consignment in good condition at our Stores & satisfactory installation and commissioning of the ordered equipment on submission of copy of DGS&D registration wherever it is applicable.
- 8) Indicate the names of the Indian reputed Organizations where you have supplied the similar equipment and may attach the satisfactory performance report of the equipment from user Organization.
- 9) a) If you have supplied identical or similar equipment to other Institutes under Ministry of Earth Sciences and Ministry of Science & Technology, the details of such supplies for the preceding three years should be given together with the prices eventually or finally paid.  
b) Based on the above information IITM will have its option to obtain details of the equipment, their performance, after sales services etc. for evaluation of the tender, directly from the concerned Labs. /Scientists etc.

- 10) **Fax / E-mail / Telegraphic / Telex tenders will not be considered.**
- 11) **All the Bank Charges inside and outside India, including opening of LC, communication, confirmation, amendments etc., if any to Beneficiary's Account only. This may please be noted.**
- 12) Details of services rendered by you as well as after-sales services offered by you are to be made clear in the tender.
- 13) The Tenderer's conditions printed on the tender or otherwise sent along with the tender shall not be binding on IITM.
- 14) The vendor should have appropriate facilities and trained personnel for supply, installation, commissioning and warranty-maintenance of the equipment to be supplied. Detailed information in this regard may be furnished.
- 15) **Delivery Period:** As time is the essence of the contract, Delivery period mentioned in the Purchase Order should be strictly adhered to.
- 16) Details regarding terms of payment including period of warranty. However we prefer to release the payment on BILL Basis (excluding Indian Agency Commission) after receipt of consignment in good condition and satisfactory installation, and commissioning thereof. Alternatively, depending upon the value and foreign exchange regulations the payment can also be considered through Sight Draft / Letter of Credit through the State Bank of India for the order value excluding the Indian Agency Commission. However, the detailed payment for Sight Draft / Letter of Credit is mentioned at clause No. **26 (B) & (C)** of this tender terms and conditions.
- 17) Supply means "Supply, Installation, Commissioning and satisfactory demonstration of the whole system and training". If any charges extra for Installation, Commissioning and training, the same should be specified in the commercial offer.
- 18) Kindly attach a copy of your latest DGS&D, New Delhi registration Certificate under the compulsory Scheme of Ministry of Finance regarding the registration of Indian Agent of foreign supplier wherever it is applicable.
- 19) a) The Tenderer is required to furnish the Permanent Account Number (PAN) Allotted by the Income Tax Department. If registered with the National Small Industries Corporation, the registration number, purpose of registration and the validity period of registration' etc. should also be provided in Technical Bid for Indian Agents.
- b) A copy of latest Income Tax Clearance Certificate from Income Tax Department (INDIA) for Indian Agents.
- 20) In case of foreign quote, the address of Principal's / Manufacturer's and their Banker's details should be furnished.
- 21) The item should be supplied with manuals and the manuals including technical / Electronic drawings / circuit diagrams should be complete in all respects to operate the system without any problem.

- 22) The import license is covered under OGL policy of 2004 - 2009.
- 23) a) In case of Indigenous Items the offer should contain the Basic Price and percentage of Excise Duty should be shown separately, since IITM, Pune is exempted from payment of Excise duty vide Govt. Notification No.10/97-Central Excise dated 15<sup>th</sup> March, 1997.
- b) IITM is exempted from payment of Custom Duty vide Govt. Notification No.51/96-Customs dated 23<sup>rd</sup> July, 1996. However as per the Govt. of India further notification No.24/2002-Customs dated 15<sup>th</sup> March, 2002 and notification No.19/2006-Customs' dt. 01.03.2006 Custom Duty is levied on all imports meant for IITM. Since the suppliers are requested to quote only on FOB basis freight, insurance and custom duty as applicable to R&D Institutions will be paid by IITM.
- 24) Tender must clearly indicate the features offered unit price, VAT tax, transport, transit insurance, installation charges. Institute cannot furnish any concessional certificate for exemption or reduction in VAT tax or any other duty / tax. The vendor should mention the price of the equipment and the duties / taxes to be paid such as customs duty / excise duty / VAT taxes etc. separately.
- 25) a) No advance can be paid.
- b) **PAYMENT:** 90%' payment shall be made by Sight Draft / an Irrevocable Letter of Credit established in favour of' the supplier through the State Bank of India, Main Branch, Camp, Pune 411 001 (India) for the order value, excluding the Agency Commission due to the Indian Agents, against the presentation of original Shipping documents. Balance 10% will be released after completion of satisfactory installation, commissioning, demonstration of the whole system, after imparting training and upon receipt of Bank Guarantee from nationalized bank for 10% of total Order value towards performance security to be valid for one year from the date of installation. However Letter of Credit/Sight Draft arrangement will be made for 100% order value.  
The Agency Commission to the Indian Agent will be paid only after successful installation, commissioning and satisfactory demonstration and acceptance of the items ordered for by the end user.
- c) For Indigenous items, 90% payment shall be made against delivery, installation, commissioning and on acceptance as per Purchase Order at site and balance 10% shall be made after receipt of performance Bank Guarantee from nationalized bank for 10% of the total order value, to be valid for entire period of warranty including extension if any. If no Bank Guarantee is given, the balance 10% will be paid after assessing, after sales service during warranty period i.e. payment after warranty period.
- 26) The prices quoted should be firm and irrevocable and not subject to any change whatsoever, even due to increase in cost of raw material components and fluctuation in the foreign exchange rates and excise duty.
- 27) The Tenderer has to state in detail the Electrical Power/UPS requirements, floor Space, head room, foundation needed and also to state whether Air-conditioned environment is needed to house the system and to run the tests. i.e. pre-installation facilities required for installation may please be intimated in the technical bid.

- 28) **INSTALLATION:** Bidder shall be responsible for installation / demonstration wherever applicable and for after sales service during the warranty and thereafter.
- 29) Acceptance tests to be prescribed later will be carried out after installation and the items will be taken over only after successful completion of the acceptance tests.
- 30) Tenders addressed to the Director, Indian Institute of Tropical Meteorology, Pune 411008 are to be submitted for each item in duplicate in double cover, under two bids system. Superscribed with Tender No. **PS/125/27/2009** for purchase of “  
\_\_\_\_\_” due on **17 June 2009**.
- 31) The tenders must be clearly written or typed without any cancellations / corrections or overwriting.
- 32) The tenders will be received in the Institute till **17 June 2009** up to **17:00 hrs.** and shall be opened on **18 June 2009** at **15:00 hrs.** in presence of the tenderers or their authorized agents who wish to be present.
- 33) IITM will not be responsible:  
a) For delayed / late quotations submitted / sent by Post / Courier etc.  
b) For submission / delivery of quotations at wrong places other than the Office of Director, IITM, Pune ,
- 34) In case of Foreign Quote, the mode of dispatch should be by Air Post Parcel/Ocean Freight/Air Freight (By Air India Freight) and on Freight to-pay basis only. The approximate dimensions of the packages and weight of consignment are to be indicated.
- 35) The makes / brand and name and address of the manufacturer, Country of Origin, Country of Shipment and currency in which rates are quoted are to be mentioned.
- 36) The Technical Bid should accompany with complete specification, Manufacturer's name, address and relevant Technical Literature/Brochures with Warranty Terms.
- 37) The payment of local currency portion shall be payable in equivalent Indian Rupees, within 30 days after the receipt of the equipment in good condition and after satisfactory installation and commissioning and demonstration.
- 38) The tender / quotation / offer submitted by you should be valid for a minimum period of **One Hundred Eighty (180) days** from the date of opening the tender.
- 39) The quotation should be only in Indian Rupees for indigenous items. In case of foreign quote, the vendors may quote their rates in, Indian Rupees as well as in Foreign Currency.
- 40) The bidder who submits the tender on behalf of their principals should produce documentary evidence in support of their authority to quote or submit proforma invoice of their principals.
- 41) Installation demonstration to be arranged by the supplier free of cost and the same is to be done within 15 days of the arrival of the equipment at site after site readiness.

- 42) IITM will not provide any accommodation/transportation for the Engineers/ Representatives for attending Installation, Commissioning and Demonstration Work. It is the absolute responsibility of the Principal Supplier/Indian Agent to make their own arrangements.
- 43) **WARRANTY/GUARANTEE:**
- (a) The equipment is to be guaranteed for trouble free performance during OMC period after installation. Supplier shall finally warrant that all the stores, equipment and components supplied under the ORDER shall be new and of the first quality according to the specifications and shall be free from the defects (even concealed fault, deficiency in the design material and workmanship). The defects, if any, during the guarantee period are to be rectified free of charge by arranging free replacement wherever necessary. Further, the technical specifications and requirements may also be verified and quoted accordingly.
- (b) Instrument should remain operational at least 90% of the time in a year and should not remain non-functional continuously for more than **5 days** in general and not more than 10 hours during most critical period Oct 1-15, 2010 during which CWG-2010 will be held. In case of non-compliance the OMC will be extended proportionately. The instrument should be commissioned, regularized and operational after the order is executed without any additional facility and cost to IITM excepting the infrastructure which includes well furnished laboratory space with 230V standard Indian power supply.
- 44) Details of onsite warranty, agency who shall maintain during warranty and undertake Operational Comprehensive Service Maintenance Contract (OMC) beyond warranty shall be given in the offer. In case of foreign quote, the Indian Agent who shall maintain during warranty and AMC / OMC beyond warranty shall be given in the Technical Offer.
- 45) The supply of spare parts is to be guaranteed at least for a period of 10 years after the supply of the equipment.
- 46) Please mention that during warranty period who will maintain system/ equipment / instrument. Indicate the name of firm, address, contact person, phone no. and fax no. etc in your technical bid.
- 47) After successful installation what will be the minimum down time of equipment/instrument in case of breakdown. If the identified firm or person fails to put the system into working condition what is the further alternative course of action suggested by you to adhere to minimum down time.
- 48) Warranty period will stand extended for a period of total downtime of the equipment.
- 49) After warranty period (post warranty) who will maintain equipment / instrument. Indicate the name of Firm, address, contact person, phone no. And fax no. etc in your technical bid.
- 50) Kindly mention the charges separately for optional item in 'K' of the Tender Document (Technical Specifications) in **commercial bid** (for post warranty period).

- 51) No sub-contracting will be allowed for installation or maintaining system/ equipment / instrument during or after warranty period.
- 52) You have to quote rates for AMC / OMC, if any, only in commercial bid.
- 53) Any upgrade in Operating System (OS) and associated other software during the warranty period should be supplied free of charge.
- 54) Discount offered should be mentioned clearly in the commercial bid only.
- 55) The Tenderers are requested to quote for Educational Institutional Price for Equipment and Software, since we are eligible for the same.
- 56) This Institute is an autonomous scientific research organization under the Ministry of Earth Sciences and is a recognized centre for studies leading to M.Sc. and Ph.D. of the University of Pune and various other Universities. As such, all possible concessions / discounts / rebates applicable for educational Institutions may be given.
- 57) You have to submit two separate bids in two separate envelopes and you may keep both the bid envelopes in an envelope for sending to us.
- One envelope will contain only the TECHNICAL SPECIFICATIONS of the indented equipment.
- Another envelope will contain only the financial bid in which price, maintenance, AMC / OMC etc. and any other information, which has financial implications, will only be given.
- The main envelope, which will contain both the bids, should be super scribed with our tender enquiry No. **PS/125/27/2009** due on **17 June 2009**.
- 59) The technical bids will be opened on the specified due date in the presence of tenderers who wish to be present & the financial bids of only those bidders will be opened whose technical bid is found suitable by us.
- 60) The Date and Time of opening for Part-II (Commercial Bid) will be intimated only to pre-qualified and technically acceptable Tenderers for the item at a later date.
- 61) a) The Earnest Money Deposit of **Rs.11,00,000/- (Rs. Eleven Lakha only)** must be paid / sent along with your technical bid in the form of a Demand Draft, Banker cheque or Bank Guarantee (from a Nationalized Bank only) drawn in favour of The Director, Indian Institute of Tropical Meteorology, Pune payable at Pune, otherwise your technical & financial bids will not be considered at all. The Earnest Money of successful bidder will be returned only after installation, commissioning, satisfactory demonstration and on acceptance of the equipment by the user Scientist / HOD as per the terms of our purchase order. If the successful bidder fails to fulfill the contractual obligations before the due date, he will forfeit the EMD.
- The Earnest Money of the unsuccessful bidder whose technical bid has not been found suitable will be returned within 15 days after receipt of Technical Committee recommendations.
- b) Those who are registered with Central Purchase Organization (e.g. DGS&D), National Small Industries Corporation or the concerned Ministry / Department need not to furnish EMD along with their bids.

- c) Though EMD has to be submitted by Demand Draft, Banker's Cheque or Bank Guarantee, we prefer to have Bank Guarantee for easy return to the bidders once a decision is taken by IITM. (Specimen of Bank Guarantee is enclosed at Annexure 'A').
- d) Tenders not accompanied with Demand Draft I Bank Guarantee towards "Earnest Money Deposit" will summarily be rejected.
- 62)a) Tenders, which are submitted without following the Two-Bid Offer System, will summarily be rejected.  
b) Unsigned Tenders will also be rejected.  
c) Part and incomplete tenders are liable to be rejected.
- 63)If the supplier fails to Supply, Install and Commission the system as per specifications mentioned in the order within the due date, the Supplier is liable to pay liquidated damages of one percent value of the Purchase Order awarded, per every week delay subject to a maximum of 10% for every week beyond the due date and such money will be deducted from any money due or which may become due to the supplier.
- 64)Goods should not be dispatched until the Vendor receives a firm order.
- 65)Please indicate page nos. on your quotation ex. If the quotation is containing 25 Pages, please indicate as 1/25, 2/25, 3/25 -----25/25.
- 66)The purpose of certain specific conditions is to get or procure best Equipment / service etc. for IITM. The opinion of Technical Committee shall be guiding factor for Technical short listing.
- 67)Firms which have already supplied similar equipment to IITM and have not completed required installation / commissioning / after sales service / warranty replacements etc. such firms offers will not be considered for further evaluation and no enquiries thereafter will be entertained.**
- 68)In the event the Manufacturer / Supplier proposes for amalgamation, acquisition or sale of its business to any firm during the contract period, the Buyer/Successor of the Principal Company are liable for execution of the contract and also fulfillment of contractual obligations i.e. supply, installation, commissioning, warranty, maintenance/replacement of spares accessories etc. with the same cost / ordered value while submitting your bid, you may confirm this condition.**
- 69)Conditional Offers will not be considered.
- 70)All disputes are subject to exclusive jurisdiction of Competent Court and Forum in Pune, India only.
- 71)The Director, Indian Institute of Tropical Meteorology, Pune 411 008, India reserves the right to accept any tender in full or in part or to reject the lowest or any or all tenders without assigning any reason.

72) In case of any dispute regarding part-shipment, non-compliance of any feature etc., The Director, Indian Institute of Tropical Meteorology, Pune will be the final authority to decide the appropriate action and it will be binding on the Vendor.

**73) Last Date and Time for receipt of Tenders: Up to 17.00 hrs. on 17 June 2009.**

**74) Date and Time of opening of Tenders: At 15:00 hrs. on 18 June 2009.**  
(Part - I Technical Bid only)

(S. R. Nirgude)  
Senior Technical Officer-II  
For Director  
**Email: nirgude@tropmet.res.in**

## BID SECURITY FORM

Whereas 1 (hereinafter called " the Bidder") has submitted its bid dated (date of submission of bid) for the supply of \_\_\_\_\_ (name and/or description of the goods)(hereinafter called "the Bid").

KNOW ALL PEOPLE by these presents that WE \_\_\_\_\_(name of bank) of (name of the country), having our registered office at (address of bank)(hereinafter called "the Bank"), are bound unto (name of Purchaser) (hereinafter called "the Purchaser") in the sum of \_\_\_\_\_ for which payment well and truly to be made to the said Purchaser, the Bank binds itself, its successors, and assigns by these presents. Seal~d with .the Common Seal of the said Bank this \_\_\_ day of 20\_\_\_\_\_  
THE CONDITIONS of this obligation are:

1. If the Bidder withdraws it's bid during the period of bid validity specified by the Bidder on the Bid Form; or
2. If the Bidder, having been notified of the acceptance of it's bid by the Purchaser during the period of bid validity:
  - a) fails or refuses to execute the; Contract Form if required; or
  - b) fails or refuses to furnish the performance security, in accordance with the Instruction to Bidders.

We undertake to pay the Purchaser up to the above amount upon receipt of its first written demand, Without the Purchaser having to substantiate its demand, provided that in its demand the Purchaser will note that the amount claimed by it is due to it , owing to the occurrence of one or both of the two conditions, specifying the occurred condition or conditions.

This guarantee shall remain in force up to one year after the period of the bid validity, and any demand in respect thereof should reach the Bank not later than the above date.

(Signature of the Bank)  
Name of Bidder.