



भारतीय उष्णदेशीय मौसम विज्ञान संस्थान
INDIAN INSTITUTE OF TROPICAL METEOROLOGY,
(पृथ्वी विज्ञान मंत्रालय का स्वायत्त संस्थान, भारत सरकार के अधीन)
(An Autonomous Institute of the Ministry of Earth Sciences, Govt. of India)

सं. / No

PS/125/75/2010/

दिनांक / Date

13 May 2011

To

Sub: Purchase / Import of "Liquid Water Content Probe (Hot Wire) 0-3 g/m³" – Qty 02 Sets.

Ref: This Institute's Enquiry Letter No.PS/125/75/2010/207-222 dated 12.04.2011.

Dear Sir,

Please refer to this Institute's above enquiry letter; it is to inform you that the last date of receipt of quotation is extended up to **14.06.2011**. You may please arrange to send the quotation positively on or before **14.06.2011 up to 12:00 hrs.** The quotation will be opened on **same day (i.e. 14.06.2011) at 15:00 hrs.** The specifications, terms and conditions would remain unchanged.

Thanking you,

Yours faithfully,


(V.R. MALLI)
Senior Technical Officer-II
For Director

तार : ट्रॉपमेट, पुणे Grams : TROPMET, PUNE

फैक्स : Fax : (020) 25893825 दूरभाष : Telephone : 25893600 / 25893675 - 81

डॉ. होमी भाभा मार्ग, पाषाण / Dr. Homi Bhabha Road, Pashan, पुणे / Pune - 411 008 (भारत / India)

**Indian Institute of Tropical Meteorology, Pashan,
Pune – 411 008 (INDIA)**

(Enclosure to Enquiry Letter No. PS/125/75/2010/ dated 13.05.2010)

**Specifications of “Liquid Water Content Probe (Hot Wire) 0-3 g/m³”–Qty 02
Sets**

- Liquid water content (LWC) measuring range 0 - 3 g/m³
- Interchangeable circuit card sensor
- Pulse width modulation power control
- 25 Hz frequency response
- Signal output 0 - 10 V
- System power: 28 VDC, 7.5 A max
- Anti-ice power: 28VDC, 8.0 A

(V. R. Mali)
Senior Technical Officer-II
For Director