

INDIAN INSTITUTE OF TROPICAL METEOROLOGY
DR. HOMI BHABHA ROAD, PASHAN, PUNE

PS/125/35/2016

15 November 2016

Sub: Minutes of the Pre-bid meeting of the "Supply and Installation of Combined Weather Sensor with Gadget" - Qty 05 Sets and "Supply and Installation of Multi-Weather Sensor" - Qty 02 Sets (as per tender document) held on 15/11/2016 at IITM, Pune.

A Pre-bid meeting relating to the "Supply and Installation of Combined Weather Sensor with Gadget" - Qty 05 Sets and "Supply and Installation of Multi-Weather Sensor" - Qty 02 Sets held on 15th November, 2016 at 11:30 hrs. at IITM, Pune.

In response to our Tender Notice No.PS/125/34/2016 and No.PS/125/35/2016, representatives of the following prospective bidders / firms / companies had attended the meeting.

- i) M/s. Dynalog India Limited, Mumbai / Pune
- ii) M/s. Environnement SA India Pvt. Ltd., Mumbai

Representative of the prospective bidders / firms / companies were asked about any suggestion, queries or technical advancement regarding tender document & Equipment to be procured. The Institute's reply to the queries raised by M/s. Dynalog India Limited, Mumbai / Pune and M/s. Environnement SA India Pvt. Ltd., Mumbai is as given below:

1. Web based software required for data logger or only data logger configuration and data download software on local PC will serve.

Reply: only data logger configuration and data download software on local PC will serve.

2. GSM modem required on data logger

Reply: NO.

3. Tender docs say only serial output of Sensors, whether analog / outputs are accepted.

Reply: Analog output are also accepted along with serial outputs

4. Please specify Storage capacity of data logger

Reply: In GB range.

5. Please specify scanning rate of data logger

Reply: 1 Hz rate and 10 Hz and also it should be user selectable scanning rate.

6. Tower for mounting sensors required? If yes, what is the height of the tower?

Reply: No.

7. Enclosure required for data logger? If yes, what is IP rating of enclosure?

Reply: Yes, IP66 standard.

8. Installation locations for all two sets.

Reply: Delhi, Hissar and Pune.

9. Revised Last date of submission of bids: 14th December 2016 upto 12:00 hrs.

Revised Opening of bids (Technical only): 14th December 2016 at 15:00 hrs.

The Pre-bid Committee suggested continuing / keeping the tender document as it is incorporating the changes detailed at Sr. No.1, 2, 3, 4, 5, 6, 7, 8 and 9 above.