

## Indian Institute of Tropical Meteorology

PS/125/22/2016

9<sup>th</sup> January 2018

### **Sub : Minutes of the Pre-bid meeting of “Purchase & Installation of Micrometeorological Sensors with Data logging System” Qty -01 Set (as per the tender document) held on 27<sup>th</sup> December at IITM, Pune**

In response to our Global Tender Notice No. PS/125/22/2018, representative of the following prospective bidder / firm / company had attended the meeting:

- 1) M/s Sutron Hydromet Systems Pvt. Ltd

At the outset Convener of the meeting welcomed the members of the committee & representative of the prospective bidders / firms / companies. He briefed the tender document, scope of supply and technical parameters of the equipment to be procured.

Representative of the prospective bidder / firm / company was asked about any suggestion, queries or technical advancement regarding tender document & Equipment to be procured. The Queries and Institute's reply to the queries are as given below:

#### **A . Queries raised by M/s. Sutron Hydromet Systems Pvt. Ltd, New Delhi and Reply by the Institute:**

- 1) Where shall you receive the AWS data from the datalogger via GSM/GPRS  
**Response:** Data will be received online in real time at the IITM Pune.
- 2) Whether the LAN Cable for the Ethernet is in the scope of the vendor or institute?  
If in scope of vendor, then what is the length? If not, then will IITM provide the cable till the tower?  
**Response:** IITM will provide LAN connection up to the tower base and then vendor has to manage the connection to the instrument.
- 3) Is battery and solar panel to be quoted by vendor  
**Response:** Battery and Solar Panel quotation not required in this bid.
- 4) Is GPRS optional or compulsory  
**Response:** GPRS is compulsory
- 5) That IITM already has the tower, what about the sensor mounting hardware  
**Response:** Vendor will have to provide all the sensor mounting accessories
- 6) Will data be downloaded manually? Do you need the option of automatic data download  
**Response:** The Institute needs both provision for data download, manual and automatic.

The committee ended with vote of thanks to the members.