

Indian Institute of Tropical Meteorology

PS/125/08/2019

16th September 2019

SUB : Minutes of the Pre-bid meeting of “ Supply and Installation of HTDMA ” Qty -01 Set (as per the tender document) held on 16th Sept.2019 at IITM, Pune

A Pre-bid meeting for purchase of Hygroscopic Tandem Differential Mobility Analyzer (HTDMA) was held at IITM, Pune on 16th September 2019 at 11.00 hrs.

In response to our Global Tender Notice No. PS/125/08/2019, no representatives of the prospective bidders had attended the meeting. This Institute had received queries from M/s Mars Bioanalytical Pvt. Ltd., Delhi vide their e-mail Dt.13.09.2019.

Technical and Commercial queries raised by the above mentioned prospective bidder have been addressed by the Committee and replies to the queries are as given below.

Queries raised by M/s. MARS Bioanalytical Pvt. Ltd, New Delhi

1) Point No. (8) Pre-impactors :- Please change the pre-Impactor cut point Size to read as 2- μ m Cut Pont Pre – Impactor.

Response: Ok, agreed.

2) Clause 1.15.7:- Please clarify if an authorized distributor of foreign manufacturer registered with MSME is exempted with Bid Security / EMD

Response: If the prime bidder is registered with MSME then only the bidder is exempted with Bid Security / EMD

3) Clause 4.5: - (i) On Site Comprehensive Warranty:- Please amend the Onsite Comprehensive Warranty as M/s BMI USA provides one year standard warranty. 2nd Year extended warranty will be quoted extra separately. The Performance Bank Guarantee will be valid for Standard One year Warranty

Response: Amendment cannot be done as 2 years warranty is needed. Regarding Performance Bank Guarantee, the Performance Guarantee should be valid for two months after the total warranty period.

4) Clause 4.6 Delivery Schedule:- Please amend the delivery to read as 20-24 weeks after receipt of Letter of Credit.

Response: Please read it is 12 weeks after receipt of Letter of Credit.